



The Abdus Salam
International Centre for Theoretical Physics



EMAS 2008

8th EMAS Regional Workshop on
**Electron Probe Microanalysis
of Materials Today**
Practical Aspects
including a session on
synchrotron-based microanalysis

19 - 22 April 2008
Adriatico Guesthouse
The Abdus Salam International Centre for Theoretical Physics
Trieste, Italy

Book of Abstracts

SYNCHROTRON MICROANALYSIS OF COMETARY PARTICLES

Laszlo Vincze^{1*}, B. Vekemans¹, G. Silversmit¹,
T. Schoonjans¹, S. Schmitz², M. Burghammer³, C. Riekel³
and F. Brenker²

1 *Ghent University, Department of Analytical Chemistry
Krijgslaan 281 (S12), BE-9000 Ghent, Belgium*

2 *Johann Wolfgang Goethe Universität
Frankfurt, Germany*

3 *European Synchrotron Radiation Facility (ESRF)
Avenue des Martyrs, P.O. Box 220, FR-38043 Grenoble, France*

* *laszlo.vincze@ugent.be*



ABSTRACT

X-ray fluorescence tomography and polycapillary based confocal XRF imaging using synchrotron radiation are among the most sensitive, non-destructive microanalytical methods which can provide potentially three-dimensional (3D), quantitative information on the elemental distributions in the probed sample volume with trace-level detection limits.

The above mentioned micro-XRF imaging methods have been applied for the non-destructive 2D/3D elemental analysis of unique comet coma micro-particles which were sampled and brought to Earth in the framework of NASA's Stardust mission. This unique and previously unseen material - captured in aerogel during the NASA mission - has been examined by a large number of research groups during the preliminary examination period using various non-destructive microanalytical techniques [1 – 3].

The examined aerogel samples typically contained a single elongated cavity (called track) produced by the impact of a single comet coma particle. We performed multiple experiments on such tracks and their terminal particles at the ESRF beamline ID13 using 2D and 3D nano X-ray fluorescence (XRF) and scanning X-ray diffraction. The XRF measurements have been carried out using two SDD detectors positioned orthogonally with respect to the incoming beam in a symmetrical arrangement relative to the sample. The nanobeam of X-rays was produced by a linear Bragg-Fresnel lens assembly which reduced the dimensions of the beam to the 170-200 nm level.

The presented quantitative micro/nano-XRF results have been obtained by the fundamental parameter and reverse Monte Carlo quantification methods, resulting in concentration distributions for the detectable elements within this unique cometary matter.

REFERENCES

- [1] D.S. Burnett (2006) *Science*, **314**, 1709.
- [2] D. Brownlee et al. (2006) *Science*, **314**, 1711.
- [3] G. Flynn et al. (2006) *Science*, **314**, 1731.